



PAR SENSOR AND PYR SOLAR RADIATION SENSOR INSTALLATION INSTRUCTIONS

METER

Preparation


Inspect and verify components are intact. Installation will require a 31.8- to 50.8-mm (1.25- to 2.00-in) diameter mounting pole and mounting bracket (sold separately). Before beginning installation, consider the surroundings and avoid obstructions.

Set up and test the system (sensors and data loggers) in a lab or office. Ensure the data loggers are using up-to-date firmware and software. Check the following locations for firmware and software updates:

ZL6 metergroup.com/zl6-support

Other metergroup.com/environment/articles/buy-browse-meter-legacy-data-loggers

Read the specifications and manual for the PAR at metergroup.com/par and the PYR at metergroup.com/pyr.



Sensor	Spectral Range	Measurement Range	Resolution
PAR	410–655 nm	0–4,000 $\mu\text{mol} \times \text{m}^{-2} \times \text{s}^{-1}$	2 $\mu\text{mol} \times \text{m}^{-2} \times \text{s}^{-1}$
PYR	360–1,120 nm	0–1,750 $\text{W} \times \text{m}^2$	1 $\text{W} \times \text{m}^2$

All products have a 30-day satisfaction guarantee.



METER

Installation

Mount the sensor with the cable pointing toward the nearest magnetic pole. For example, in the Northern Hemisphere, point the cable toward the North Pole. Make sure that the domed sensor body is kept horizontal. Use the leveling plate to ensure the sensor is level. Plug the sensor cable into the data logger. Use data logger software to apply appropriate settings to the sensors plugged into each data logger port. Gather and secure the cables to the mounting pole in one or more places to relieve strain on the connections.

METER Group, Inc. USA

2365 NE Hopkins Court, Pullman, WA 99163

T +1.509.332.2756 F +1.509.332.5158

E info@metergroup.com W metergroup.com

North America

E support.environment@metergroup.com

T +1.509.332.5600

Europe

E support.europe@metergroup.com

T +49 89 12 66 52 0